



Certificate of Conformity No 116-03111

Translation

<i>Object</i>	Luminaire for medicinal purpose
<i>Type</i>	CHROM-NT, serial number 12160051 / CH/100/NT Dimensions (L x W x H) : 1537 mm x 635 mm x 110 mm
<i>Ballast</i>	Integrated
<i>Applicant</i>	ZENIUM S.A. Rue Edouard Michelin 100 F-54 710 Fléville Devant Nancy
<i>Requirements</i>	Conformity of the chromaticity in respect to the CIE standard illuminant D ₆₅ , according to BS 950-1:1967.
<i>Confirmation</i>	According to the Measurement Report METAS Nr. 116-03110 from 18 th of November ²⁰¹⁶ the luminaire CHROM-NT complies with the requirements of the chromaticity in respect to the CIE standard illuminant D ₆₅ according to BS 950-1:1967.
<i>Date of Examination</i>	17.11.2016

3003 Bern-Wabern, 19th of January 2017

For the Examination

Sector Length, Optics and Time


Beat Imhof


Dr Rudolf Thalmann, Head of Sector